



## MORTAR PENETROMETER



**Model No.: SET-1111**

Standards: ASTM C-403

Specification: It is used for finding out the rate of hardening of mortar sieved from concrete spring and a stem graduated from 0-70 kg x 1 kg. Six interchangeable penetration needles of areas 645, 323, 65 32 and 16 mm sq. is provided . The penetration resistance is measured by the force exerted to penetrate the mortar by 25 mm and is indicated by a sliding ring on the stem, which is graduated. Needle shanks are marked at every 12.5 mm. Supplied complete in a wooden carrying case.

**SE-TEST Lab Instruments (I) Pvt. Ltd.**

**Address & Contact:** Plot No.: 19/261, Basti Sarai Rohilla, Near Shastri Nagar Metro Station, New Delhi-110035

**Ph.:** 011-2365-7121 **Mob.:** 91-9891445495, 8448366515 **E-mail:** setestindia@gmail.com **Website:** www.setestindia.com